

Agilent Feature Extraction Software Version 10.5.1.1 (FE 10.5.1)

(c) Copyright 2008, 2015 Agilent Technologies

For Research Use Only. Not for use in diagnostic procedures.

G4460-60004 Feature Extraction Software Release Notes

Product:

G2566AA - Feature Extraction provided with the Agilent Scanner

G4460AA - G4462AA - Feature Extraction Commercial License

G4463AA - G4465AA - Feature Extraction Academic License

This release note contains:

- * Changes in FE 10.5.1.1
- * Bug fixes to known issues in 10.1.1.1
- * Installation Notes for Feature Extraction Software 10.5.1.1
- * Known issues in Feature Extraction Software 10.5.1.1
- * Fixes to known issues in Feature Extraction Software 10.1.1.1 and 9.x

This product includes software developed by the Apache Software Foundation
<http://www.apache.org/>.

Changes in FE 10.5.1.1

1. The software provides improved grid finding for the new Agilent 30-micron feature size arrays. These grid finding changes are completely independent of 65-micron feature size arrays and it is possible to create protocols that give identical grid finding results to FE 9.5.3 for 65-micron feature size arrays.
2. The software now supports eArray integration for automatic download of missing design / grid template files, protocols and QC metrics sets from the eArray server.
3. Tiff image files in a project no longer require grid templates and/or protocols to be attached for an extraction to run. The software will automatically download the proper grid templates and automatically link the proper protocol.
4. QC metric sets are associated with FE protocols and are no longer assigned to extraction sets in a project. Each extraction will use the metric set associated with the protocol in question.
5. All FE protocols have a new protocol step, 'Optimize Grid Fit'. This protocol step adjusts the grid to improve the overall registration of the grid

on the image. This will improve spot finding for 65-micron feature size arrays and still allow users wishing to regress against previous versions of FE to turn this step off.

6. The software contains a new option available when viewing shape files. The option allows users to view only the pixels used to calculate the feature signals. This shows the affect graphically of using cookies and pixel outlier rejection to calculate the signals of each spot. The option is available on the View Menu under extraction results labeled "view pixels used".
7. When saving images, the software now allows users to save in bmp or png format as well as jpeg format.
8. The FE 10.5 installer provides a mechanism for updating FE 10.x installations leaving the previous database intact. NOTE: that FE 9.x users wishing to install FE 10.5 on the same computer as FE 9.x will need to uninstall FE 9.x
9. Axon 4000A and 4000B images cannot be extracted with FE10.5. Customers requiring this ability should use FE9.5.3
10. The CGH QC Report has been streamlined and includes the sample name field entered through the sample management application (to be released with CGH analytics in Feb 09).
11. The FENoWindows implementation has been made more robust and generates an improved XML output file that contains all error messages and all stats generated for a given array in FE. The status messages have been segregated by slide and by array so as to provide an easier mechanism for those wishing to script FE in their informatics pipeline.
12. The QC reports, txt, and XML outputs support a new field, sample name, that will be entered via the sample management application (to be released in February 2009).

=====
Bug fixes to know issues in 10.1.1.1
=====

1. In the FE 10.2 prerelease version, QC reports for the 30 micron feature size 2-pack arrays fail to generate QC reports. This issue has been partially fixed. The FE 10.5 release generates by default a new streamlined CGH QC report. This new QC report generates in all cases tested (PVCS 2019).

2. When viewing images using the Feature Extraction GUI, 5 micron scans take more memory than in FE 9.5.3 (SCR 1874).
3. Sometimes when saving the image as a JPEG generates no JPEG and no error message is generated. The workaround to this issue is to restart the software and try saving the image as a JPEG again (SCR 1920).
4. In order to speed up the Viewing of shape files, the software now by default turns on the option of viewing outliers only. This greatly improves performance when loading the shape file. The users will need to crop a region of the image and then turn off view outliers only under the Extraction Results menu in order to view all shapes (PVCS 1942).
5. Some designs could not be loaded into FE 10.1.1 or earlier release because the annotation contained in the designs were too long for the FE database. This has been fixed in FE 10.2 PVCS 1943).
6. When working with dye norm lists in FE 10, only external dye norm would work correctly. Dye norm lists attached directly to a grid template could not be loaded into the FE database. This has been partially corrected but the software still does not allow over 50000 probes to be loaded and attached to a grid template (PVCS 1948).
7. In FE 10.1, using dye normalization lists in Compact mode caused FE to stop not completing the processing of an array. The software would error with 'low level runtime error'. This issue is fixed in FE 10.2 (PVCS 1949).
8. In the FE 10.2 pre-release the new option allowing users to view only pixels used in the signal calculation, is limited to those features that are completely contained inside the cropped region. Any feature only partially inside the cropped region will have all of its pixels turned off (PVCS 1952).
9. The grid geometry dialog box did not accept enough digits of precision to modify the spot spacing the new 1 Million feature arrays. The new dialog allows for more precise entries (PVCS 1953).
10. After upgrading FE, if the user may see 2 FE shortcuts. This happens if the reboot option is not chosen. The old shortcut has been removed but the desktop is not reflecting this. Having windows refresh the desktop fixes this issue (PVCS 1965)
11. When generating a MAGE-ML file using FE 10 can cause the MAGE-ML file to have slightly incorrect information. More specifically, the laser power value, unused by the scanner, is set to a garbage value. The FE 10 software writes the laser power value out into MAGE-ML output file and can write an un-initialized

value causing the Rosetta Resolver to fail to load those effected MAGE-ML files. This has been address in FE 10.2/FE10.5 (PVCS 1970).

12. Given that we automatically assign protocols, this bug should not occur: Using the wrong protocol to extract a high density or ultra high density arrays can cause Feature Extraction to process for a very long time. This will happen if you attempt to use the CGH protocol on arrays with large numbers of negative control features (SCR 1902).
13. Sometimes opening perfectly valid scans in the FE GUI for image viewing fails with an error message like 'An invalid file handle was associated with file.tif'. Attempting to open the scan again will succeed (SCR 1788).
14. When installing FE the QC Metric set browser is not shown by default. The user must go to the view QC Metric Set Brower menu in order to get it to display (SCR 1674).

Installation Notes for FE 10.5.1.1

1. Feature Extraction can either be installed from CDROM or by downloading it from the web site <http://www.agilent.com/chem/fe>.
2. The Feature Extraction 10.5 is an upgrade release of FE 10.1 or FE 10.5. If FE 10.1 or FE 10.2 is installed then the FE 10.5 will upgrade your system without disturbing the grid templates, custom protocols or custom metric sets in your FE database. If the prior version of FE is 9.x or earlier the Feature Extraction must be uninstalled before installing Feature Extraction 10.5.1.1. To remove the previous version, go to the add/remove programs option under the control panel and remove "Feature Extraction".
3. The database in Feature Extraction 10 has changed and is not compatible with previous releases. If you use any special protocols or metric sets they must be exported from the current software before uninstalling. If you use the QC chart tool you should back up the QC chart database before installing. Please see the FE_10.5_Installation guide for more details.
4. Recommended PC configurations for FE 10.5.1.1 are using a Dual Core X86 (please see 5 below) based Processor with 2+ GB of RAM and at least 40 GB of hard disk space. Agilent recommends using 4 GB of Virtual Memory.
5. Agilent has not tested FE on Itanium Processor architectures (IA-64). As such, Agilent does not support Feature Extraction on Itanium based windows machines.

6. The supported Operating Systems are Windows XP Professional SP2 (please see 7 below), Windows Server 2003 SP1 (please see 7 below), Vista Business Edition (please see 7 and 8 below), Vista 64-bit Business Edition (please see 7, 8, and 9 below), and Windows XP 64 (please see 7 and 9 below).
7. In order to install FE the user must be a member of the administrator group. Feature Extraction needs to create directories under the "Program Files" system folder and must be able to access and edit the system registry in order to ensure correct installation.
8. On the Vista Operating System both 32 and 64 bit versions, the user must disable User Access Control before installing FE. Not disabling User Access Control (UAC) could cause the software to fail to install as it will not be able to access required system registry to ensure correct installation.
9. Current PDF QC report generation does not work in Vista 64 and XP 64. Agilent is working to solve this issue and will have a patch to FE to enable this support when appropriate drivers are available to us.
10. After uninstalling the previous version of Feature Extraction, a reboot may be required. If the reboot is requested by the installer and it is not performed then installation of the new version of FE may fail.
11. After the full installation is complete, the installer may ask to reboot the system. The reboot is required and if not done may lead to the software working improperly.
12. FE 10.5.1.1 has a new feature that allows users to run FE with their choice of databases: either SQLExpress
<http://www.microsoft.com/sql/editions/express/default.mspx> or MySQL
<http://www.mysql.com/>. Feature Extraction uses ODBC to connect to the database but we have built schema for and specifically tested against MySQL. This third party database support was specifically added for those users intending to process large number of arrays with a large number of unique designs. The SQLExpress database shipped with FE has limits on the number of designs it can hold. MySQL is an inexpensive database that does not have such limitations. Here we list some tips of installing and using FE with MySQL.
 - a. FE 10.5 uses ODBC to interact with the database, so users installing MySQL need to install the MySQL Server
<http://dev.mysql.com/downloads/mysql/5.0.html> and the Connector/ODBC - MySQL ODBC driver
<http://dev.mysql.com/downloads/connector/odbc/3.51.html>. FE 10.5 has been tested against mysql 5.0.51 on windows XP and mysql 5.0.45 on windows Vista. The ODBC tested ODBC driver is Connector/ODBC 3.51.

- b. If using MySQL, the user must install the database before installing FE. If you wish to switch to MySQL, then Feature Extraction must be uninstalled and MySQL installed before reinstalling Feature Extraction.
- c. Currently Agilent supports using Feature Extraction and a MySQL Server on the same PC. Having FE access a database on a remote machine is not supported.
- d. When installing MySQL, please make sure to enable root access from remote machines. This option appears on the MySQL Server instance configuration window when setting the security options. The Feature Extraction software connects via TCP/IP and requires that remote access be enabled.
- e. Also, when installing MySQL, please make sure to include the bin directory in windows PATH. This option is settable in the MySQL Server Instance Configuration Screen. This option is required if you wish to use Feature Extraction's Backup and Restore Utility to manage your database backups.
- f. Please contact Agilent Technical Support if you have more questions or problems.

Known Issues in Feature Extraction Software 10.5.1.1 and earlier

- Show QC chart button does not enabled when pdf QC reports and run charts cannot be generated. This means that users will not be able to use the QC chart button on XP/Vista 64 bit (PVCS 2120).
- If a user attempts to import a grid template during an extraction run, the grid template may fail to properly import into the software. The work around is to only import grid templates when a project is not running. (PVCS 2112).
- When importing text data into the QC Chart tool generated with the new CGH protocol supporting and generating streamlined CGH QC reports, the QC Report type column is blank in the QC Chart Database. The QC Report Type column is correctly loaded for all other types of QC Reports (PVCS 2109).
- In very rare instances, while running an extraction that fails to properly extract, some messages related to the extraction do not appear in the extraction summary report. The FE extraction fails with a low level runtime error (PVCS 2099).

- The user cannot abort the importation of a protocol. The software will hang and will need to be killed using the task manager (PVCS 2094).
- Protocols can be saved with a blank protocol name, but such a protocol cannot be used in the software. Adding the blank protocol to an extraction in a project will cause that project to fail to start (PVCS 2093).
- Attempt to abort an extraction on Vista 64 will cause the extraction to stop ungracefully with a low level runtime error (PVCS 2092).
- Sometimes QC Charts do not get generated at the end of a project even though all the extractions in the project use the same protocol and hence the same metric set (PVCS 2077).
- The eArray automatic online protocol update can fail if either the protocol or the metric associated with that protocol are missing (but not both) when the update occurs. It is possible to make this happen because even read only protocols and metric sets can be removed using FeNoWindows. The worst case is if the metric set is missing from FE, then the eArray will fail to update the protocol due to a missing metric set (PVCS 2075).
- Sometimes in rare instances, the switch to configure mode button will not be enabled after the extraction is complete. The extraction will have to be closed and re-opened to enable to config / run mode toggle (PVCS 2069).
- If the user attempts to remove FE via the add/remove programs while the FE instance is open, the FE un-installation fails. The user will need to run the installer to repair FE and try the un-install again with the FE instance closed (PVCS 2057).
- The grid template that is currently in use during an extraction, can be removed causing the extraction to fail (PVCS 2042).
- The eArray configuration setting “Automate FE extraction by automatically downloading design files from eArray” can cause unwanted behavior if the user wishes to extract an image using a grid template with an AMADID doesn’t match the image’s AMADID. Disabling the eArray configuration setting temporarily causes FE to behave correctly in this case (PVCS 2040).
- Attempting to calculate spot size and centroids in manual grid mode using a high resolution scan of a third party array will cause FE to crash (PVCS 2032).
- Sometimes the software will generate a memory error when try to load shape (.shp) files on high resolution scans. The work around is to close the software and restart it. Then the image and the shape file will be viewable (PVCS 2026).

- In rare instances the QC report can fail to generate. This is true when 30u feature size 2-pack arrays are used in FE, using a CGH protocol that generates an old style CGH QC report. For the 2-pack 30-micron feature size arrays the new streamlined CGH QC report must be used (PVCS 2019).
- It is not recommended to run projects containing multiple extractions directly through FeNoWindows. Projects containing multiple 30-micron feature size 1 million feature array will run out of memory if run directly using FeNoWindows. The work around is to either use the FE GUI to run these projects or to break up the project into multiple projects each containing 1 extraction (PVCS 2016).
- Do not remove the DBConnectInfo.ini file from the FE installation folder. If that file is not available the software cannot be removed via the install or Add/Remove programs (PVCS 1944).
- When viewing shape files, feature outliers are not visible until the image is zoomed or cropped (which effectively zooms) (PVCS 1955).
- Viewing the scan properties can cause the image to appear distorted. Minimizing followed by reopening the image will correct the issue (PVCS 1959).
- If the user adds a grid template to the database while in manual grid mode, then certain features of third party manual grid mode can become disabled. Specifically, if the spot mode will become disabled (PVCS 1969).
- Save histogram or line plot value for single channel scan not working (PVCS 1972).
- Sometimes QC Charts are not generated after an extraction run completes. When this happens, the software does not generate a pdf of the run charts of the selected metrics and an error message will appear run summary report. Once receiving the error message, the user will not be able to generate the chart via the menu. The only workaround is to quit and restart the FE software, followed by running the project again (SCR 1925).
- Scans of 2, 3 and even 5 micron resolution using full sized scan regions are quite large creating memory issues for the software. In order to address memory and performance issues the following restrictions are true about the imaging.
 - New format scans (e.g. 2 micron, 3 micron or 20 bit scans) cannot be flipped or rotated. 5 and 10 micron 16 bit scans are in the same format as before and can be flipped and rotated.
 - The new view window feature (with Ctrl-N as the shortcut) that allows users to open one channel of the scan will not work for new format scans.

- When cropping a new format scan to view the image close up, what FE refers to as high fidelity mode, zoom out is disabled below 200%.
- When writing creating a grid file of Agilent arrays some annotation columns may be lost. Currently only the primary annotation columns are certain to make it into the grid file. Those annotation columns are Probe Name, Systematic Name, and Gene Name. Other annotation fields are not certain to be output. There is no workaround to this issue (SCR 1917).
- For large grid templates using the dye normalization list editor can take a very long time. This is a function of the number of probes that have to read from the database and loaded into the editor. It is possible to create the lists outside of FE and load them into the software without using the list editor. Please consult the FE manual for details (PVCS 1913).
- Blank fields, namely blank annotation fields above and beyond the primary annotation fields of Probe Name, Systematic Name, Gene Name, used inside of the GEML xml design files will fail to load into the MySQL database. The MSSQL database is unaffected by blank annotation fields. The workaround is to load minimally annotated designs or to get updated designs via Agilent support (SCRs 1912 and 1853).
- Sometimes FE is unable to a create grid file for manual gridding when using third party arrays. This will happen when attempting to create a grid file from a tab text file or GAL file and either not enough information exists about the array geometry or the geometry information specified is wrong leaving the software to guess what the correct geometry is. The workaround to this is for the use to manually measure the geometry in pixels on the image before starting grid mode. The user must measure the spacing between spots in both the horizontal and vertical directions (measured spot center to spot center), the spacing between sub grids in both the horizontal and vertical directions (measured from the last spot in the first sub grid to the first spot in the next sub grid), and the diameter of each spot. Typing these values into the software if they are incorrect fixes the issue (SCR 1908).
- Data exported to Excel from the Histogram plots or line plots has the first value corrupted. All other data values after the first are correct. This only happens on Windows Vista Operating System but may be related to the version of Office being used. Under windows XP with office 2003 all data values exported to excel from the plots are correct (SCR 1884).
- Changing passwords in the database server will break the database connection for FE. For MySQL the root server password provided when FE is installed must be maintained. For MSSQL/SQLExpress the password must be "#Welcome\$" (SCR 1859).

- Under Windows Vista, FE users may not have permission to write to certain directories or overwrite certain files. This will cause FE to fail to load grid templates, protocols and/or metric sets. Also it will cause FE to fail to process extractions. The solution to this is to make sure the User Access Controls are turned off and ensure that C:\Program Files\Agilent\MicroArray\FE\Temp is writable and empty (SCR 1836).
- It is possible using the command line FE to remove a grid template while the software is processing extractions. This is not recommended as it could cause the project to fail if that grid template is used (PVCS 1826).
- Ultra high density Agilent arrays, meaning Agilent arrays containing approximately 976,000 features, scanned at 5 micron scan resolution cannot be reliably processed using Feature Extraction 10.1.1. Memory issues can surface while processing, automatic grid registration is significantly less reliable and Manual grid registration is difficult. We strongly recommend scanning these arrays at 3 micron resolution or higher (SCR 1775).
- Applying many different image manipulations including Flip and Rotate on 5 micron 16-bit scans may cause FE to run out of memory and yield the message "Failed to load bitmap". The workaround is to close the FE software re-open it and try the operation again (SCR 1874, 1768, and 1766).
- Protocols that are loaded into the database by name are specifically not case sensitive but protocols specified in the project are case sensitive. Please keep all protocol names the same case (e.g. stick with capital letters) (SCR 1699).
- When cropping an image using the new crop image dialog, the file name is appended with 1_1 although "Cropped multiplex" is not selected (SCR 1679).
- When running an XDR extraction project, the summary report can give grid placement error '... The spot centers are shifted relative to their nominal grid.' But the QC report says the grid is normal. This is an indication that XDR failed to properly extract the low PMT scan and there may be an issue with the image registration (SCR 1675).
- When using FE 10.1.1.1 with PDF QC reports and charts, sometimes a blank QC Chart is generated after extraction. Rerunning the project again will generate a corrected QC Chart. This is directly related to having Internet Explorer (IE) open when running Feature Extraction. We recommend not using IE while running FE (SCR 1667).
- When a two-channel tiff file is split into two single-channel tiff files (one for Red and one for Green) from the Agilent Microarray Scanner, On-Time project treats the split tiff files as two separate single-channel files (SCR 282).

- Protocol, DyeNormList, or GridTemplate cannot have special characters in the name or description. The character that is sure to break is ""(SCR 652 & 657).
- FE is not designed to support concurrent users. The software will only allow one user to use FE at a time. If another user is logged into the same machine, only one will be allowed to run FE.
- You can not open a .tif image from a CDROM or DVD. Copy the image to a hard disk or network share.

=====

Known issues in Previous Versions fixed in 10.1.1.1

=====

Fixed in FE 10.1.1

- 1 color Axon scans cannot be opened or processed in FE (SCR 1740).
- FE allows extraction to start without selecting the output folder. If the user opens a new standard project and in the project properties tab on the local file folder selects "same as image" as false and doesn't enter a results folder FE will crash after processing the image or images (SCR 1739).
- Running a legacy background subtraction algorithm called "Global Background Adjust" fails for multipack arrays (SCR 1732).
- Installer fails on Vista 64. SQL Express fails to install due to the path to the database being too long. The FE installer now correctly constructs a path to support the SQL Express database installation (SCR 1728).
- SQL Express fails to install properly on systems without Windows installer 3.1 installed and/or MDAC version 2.8 installed (SCR 1727).
- "FTP send PDF only" option does not work in project settings. Attempting to ftp PDF QC Reports yields the message "FTP of QC Report is not supported in 9.1" and the reports are not transferred (SCR 1726).
- Auto estimate additive error not possible for 3rd party arrays (SCR 1721).
- Spatial distribution plots incorrect drawn. For 244k CGH arrays the spatial distribution of significantly positive and significantly negative probes on the QC report is incorrect. The top half of the plot contains significantly more points than the bottom half. This is an issue with the plot and not the data generated by FE (SCR 1706).

- When running a project that contains XDR extractions, sometimes the project will fail with a low-level runtime error. This is because FE failed to allocate memory to read the low-pmt scan. This is a workaround to a more serious issue part of FE 9.1 and 9.5 causing the generation of bad data. This issue is very infrequent but if it does occur you need to exit the application and restart before running your project again (SCR 1705).
- Feature Extraction 9.5.3 can run with multiple threads (FE 10 runs only 1 thread). With the high density Agilent arrays (scanned at 5 microns) it is possible for a thread to stop running due to lack of memory. When this occurs the extraction being run in that thread re-enters the queue and gets processed by another thread. When this occurs, it is best to let the project complete successfully and then close and re-open the Feature Extraction SW before running additional projects. If enough projects are run without closing the Feature Extraction SW, occasionally even the final thread will fail to complete due to lack of memory or the application will stop responding (SCR 1489).

Fixed in 9.5.3

- Loading multi-file tiffs can assign the wrong channel. For example pick any 2-color tiff (GenePix or Agilent) spread across 2 files. Opening the Green channel file first will load the green image into channel 0. Opening the red channel tiff first will load the red image into channel 0. If the green channel is channel 0 and the image is processed then the channels will be flipped in the output file (SCR 1716).
- The use of Dye Normalization lists is broken for multipack array processing. If the user attempts to use a dye normalization list with a multipack array the software will run out of memory or encounter a low level runtime error (SCR 1704).
- Multiplicative Detrending runs on the wrong class of controls causing the algorithm so consider very few true replicates (in the case of custom designed Agilent microarrays). The effects of this can be quite different than what we would expect (SCR 1701).
- For 1-Color scans, Image properties dialog displays incorrect information on Scanner offset section of the info dialog (SCR 1670).
- MAGE-ML outputs for arrays processed with 1-color protocols and 2-color protocols which calculate the error model directly after background subtraction do not include the error model parameters, relying on Rosetta Resolver's defaults for Multiplicative and Additive Error terms. The 2-color protocols in question here were introduced with FE version 9.1 of the software

and have '91' in their name. This leads to large differences in significance calls between the Feature View of a single array and any combined view in Resolver (Reporter view, Gene view, any multiple array view). 2-color protocols which calculate the error model after the ratios are calculated, correctly output Agilent's error model parameters into the MAGE-ML file for use in Resolver and do not display this problem (SCR 1654).

- The software generated invalid MAGE-ML when a security group is not specified in the project. The security group and owner were meant to be optional so the software should have generated valid MAGE-ML without adding this (SCR 1634).
- If a user chooses to output grid file from a multi pack array, then uses that grid file to extract, the results are wrong (spot find shows all sub-arrays pointing to 1st sub-array position). Stats for sub-array 2, 3, 4 are identical (SCR 1631).
- The BGUsed column of FE is not correctly calculated when additive detrending is applied with another background method (SCR 1629).
- When cropping an Image, the image crop and save doesn't save the area cropped by the user. The software saves a different area not related to the cropped area (SCR 1618).
- Background Subtraction: The Global Background adjust is not calculated correctly. The adjustment appears to be subtracting double the calculated value. Global background adjust is turned off by default for all Agilent recommended protocols and we do not currently recommend its use (SCR 1610).

Fixed in 9.5.1

- Negative control standard deviation as calculated is not robust. This can be seen with CGH arrays that have "hot" negative control features. These features are artificially and locally high due to foreign substances on the array surface and are therefore candidates to be rejected as outliers. Normally FE would take care of this via the population outlier rejection but for CGH arrays the populations of negative controls are often smaller than minimum number of features needed for population outlier rejection. This can negatively affect downstream calculations (SCR 1616).
- European users of FE with annual licenses can get the message 'license will expire in -38952 days'. This is because we don't correctly recognize the European date format (SCR 1615).

- The Run chart key of run charts viewed after the completion of an FE batch do not display pack information about the barcodes in the run chart making it impossible to distinguish which array a chart value refers to (SCR 1614).
- The local background subtraction option is completely disabled in FE 9.1 and FE 9.1.3. Enabling local background subtraction in the protocol and then processing arrays with that protocol yield no difference in background subtracted signals due to local background (SCR 1612).
- Image manipulation functions, flip and rotate image are not working after loading only green channel image (SCR 1599).
- Processing of split and rotated tiff images through FE 9.1 and FE 9.1.3 is not possible, but was previously possible in FE 8.5. Attempting to process split and rotated tiff images yields an error from the software.
- On the QC Report when using a metric set, metric without thresholds were being evaluated and always showing a red 'evaluate' designation. Since there is no threshold, they should show as normal (SCR 1591).
- QC Report doesn't correctly position *** for SpikeIn Table (SCR 1588).
- Multipack arrays manually gridded then processed through FE yield strange results, the 1st array is OK all arrays after this array have double entries in the SpikeIn table. The SpikeIn plot is also affected (SCR 1584).
- By selecting compact text mode as an output type along with MAGE-ML generation yields a different and incorrect MAGE-ML output file. Selecting full text mode yields a correct MAGE-ML file. Compact tab text output should have no effect on MAGE-ML whatsoever (SCR 1571).
- The Red Multiplicative Error Value is incorrectly set for the GE2_NonAgilent_91 protocol (SCR 1570).
- The row and column information is incorrectly output on 3rd party arrays. The issue affects the QC report as well causing the outlier plots and the up and down regulated plots to become scrambled (SCR 1555).
- In the 1-color QC Report, the spike in detection limit in the "Spike-In Concentration-Response Statistics" section incorrectly shows -1, instead of real number (SCR 1484).
- Feature Extraction compact output does not include 2 parameter fields that are very useful and desirable in the QC Chart tool. The columns are FeatureExtractor_ColorMode and FeatureExtractor_QCReportType (SCR 1648).

- The percent GeneNonUniform Outliers in the stats table and in the QC Report is incorrectly calculated for two color microarrays. The percentage is about 1/2 of what it should be.
- Feature Extraction can now be installed on windows Vista OS and on 64-bit operating systems for X64 architecture machines.